

# Hellenic Accreditation System



Annex F2/2 to the Certificate No.983

## SCOPE of ACCREDITATION

Of the

**Calibration Laboratory**

**CALIBRATION & TESTING GLOBETECH LABORATORIES S.A.**

Measurand / Calibration item	Range of measurement	Uncertainty* (Confidence level 95%)	Remarks
Force Measurements			
Compression testing machines	200kN... 2000 kN	0,15 %	Machines of class1. Calibration according to standard ISO 7500-1: 2015. Calibration can be performed on-site.
Torque wrenches	0 Nm ... 500 Nm	0,6 %	Method code <b>O1002</b> Using torque measurement system.
Dimensional Measurements			
Calipers, analog / digital	0 mm ...300 mm	0,6xd d: instrument division in $\mu\text{m}$ $d \geq 10 \mu\text{m}$	Calibration according to VDI2618- 9.1: 2006

\* Where uncertainty is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.

Site of assessment: **Permanent laboratory premises, Kyprion Agoniston 46, 12 462 Chaidari, Athens, Greece**

Approved Signatories: **M. Avraam, N. Stathatos**

This scope of Accreditation replaces the previous one dated 08.06.2015.

The Accreditation Certificate No. 983, to ELOT EN ISO/IEC 17025:2005, is valid until 7.06.2019.

Athens, 02.02.2018

Konstantinos Voutsinas

*Managing Director, ESYD*