

Hellenic Accreditation System



Annex F2/1 to the Certificate No. 983

SCOPE of ACCREDITATION

Of the

Calibration Laboratory

CALIBRATION & TESTING GLOBETECH LABORATORIES S.A.

Measurand / Calibration item	Range of measurement	Uncertainty* (Confidence level 95%)	Remarks
Force Measurements			
Compression testing machines	200kN... 2000 kN	0,15 %	Machines of class 1. Calibration according to standard ISO 7500-1: 2004. Calibration can be performed on-site.
Torque wrenches	0 Nm ... 500 Nm	0,6 %	Method code O1002 Using torque measurement system.
Dimensional Measurements			
Calipers, analog / digital	0 mm ... 300 mm	0,6xd d: instrument division in μm $d \geq 10 \mu\text{m}$	Calibration according to VDI 2618-9.1: 2006
Depth calipers, analog / digital	0 mm ... 300 mm	0,6xd d: instrument division in μm $d \geq 10 \mu\text{m}$	Calibration according to VDI 2618-9.1: 2006

* Where uncertainty is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.

Site of assessment: **Permanent laboratory premises, Kyprion Agoniston 46, 12 462 Chaidari, Athens, Greece**

Approved Signatories: **M. Avraam, N. Stathatos**

The Accreditation Certificate No. 983, to ELOT EN ISO/IEC 17025:2005, is valid until 7.06.2019.

Athens, 8th June 2015

Ioannis Sitaras

Director of the Laboratories Accreditation Division